



Serving Scientific Community  
Since 1985



**TECHNO PHARMCHEM**

An ISO 9001 : 2008 Certified Company

Registered Off. : 101-A, Deep Enclave, Pocket 'D', Ashok Vihar, Phase-III, Delhi-110 052 (India)  
Corporate Off. : 152, Vardhman City Centre, Near Shakti Nagar Underbridge, Delhi-110052 (India)  
Works : Plot No.1022, Modern Industrial Estate, Bahadurgarh-124507, Haryana (India)  
Telephone : 0091 - 11 - 23646422 E-mail : tecpharm@gmail.com  
Website : www.technopharmchem.com

### CERTIFICATE OF ANALYSIS

Name of the Sample : Oxalic Acid 0.05mol/L (0.1N) for 500ml solution  
Batch No. : 224446  
Date of Mfg : Jan 2021  
Date of Exp : Dec 2023  
Register No. : 2021-22  
Qty of Sample : 1 Ampoule  
Date of Analysis : Jan 2021

S. No	Test Parameters	Observed Values	Standard Values
1.	Description	Passes	A Clear and colorless Solution
2.	Concentration after dilution to 500 mL c(C <sub>2</sub> H <sub>2</sub> O <sub>4</sub> .2H <sub>2</sub> O)	0.05 mole/l	0.05 mole/l
3.	Amount of substance concentration c(C <sub>2</sub> H <sub>2</sub> O <sub>4</sub> )	0.0501 mole/L	0.05 mole/L ± 0.2%
4.	Titer at 27°C	1.000	1.000

Chief Chemist