



Serving Scientific Community
Since 1985



TECHNO PHARMCHEM

An ISO 9001 : 2008 Certified Company

Registered Off. : 101-A, Deep Enclave, Pocket "D", Ashok Vihar, Phase-III, Delhi-110 052 (India)
Corporate Off. : 152, Vardhman City Centre, Near Shakti Nagar Underbridge, Delhi-110052 (India)
Works : Plot No.1022, Modern Industrial Estate, Bahadurgarh-124507, Haryana (India)
Telephone : 0091 - 11 - 23646422 E-mail : tecpharm@gmail.com
Website : www.technopharmchem.com

CERTIFICATE OF ANALYSIS

Name of the Sample : **Buffer Concentrate for 500ml Buffer Solution pH 7.00 +/-0.02 at 20°C**
Batch No. : **225900**
Date of Mfg : **Aug 2022**
Date of Exp : **July 2025**
Register No. : **2022-23**
Qty of Sample : **1 Ampoule**
Date of Analysis : **Aug 2022**

S. No	Test Parameters	Observed Values	Standard Values
1.	Appearance	Passes	A Clear colorless solution
2.	pH (after dilution of the concentrate with distilled water to 500ml at 20°C	7.01	7.0 +/- 0.02

Chief Chemist