



Serving Scientific Community  
Since 1985



**TECHNO PHARMCHEM**

An ISO 9001 : 2008 Certified Company

Registered Off. : 101-A, Deep Enclave, Pocket 'D', Ashok Vihar, Phase-III, Delhi-110 052 (India)  
Corporate Off. : 152, Vardhman City Centre, Near Shakti Nagar Underbridge, Delhi-110052 (India)  
Works : Plot No.1022, Modern Industrial Estate, Bahadurgarh-124507, Haryana (India)  
Telephone : 0091 - 11 - 23646422 E-mail : tecpharm@gmail.com  
Website : www.technopharmchem.com

### CERTIFICATE OF ANALYSIS

**Name of the Sample** : **Mac Conkey Agar w/o CV w/0.15% Bile Salt**  
**Batch No.** : **226822**  
**Date of Mfg** : **Aug 2023**  
**Date of Exp** : **July 2026**  
**Register No.** : **2023-24**  
**Qty of Sample** : **500gm**  
**Date of Analysis** : **Aug 2023**

S. No	Test Parameters	Observed Values	Standard Values
1.	Appearance	Passes	Light yellow to pink Homogeneous free flowing powder
2.	Gelling	Passes	Firm, comparable with 1.2% agar gel
3.	Prepared medium	Passes	Orange Red colored Clear to slightly opalescent gel forms in petri plates.
4.	Reaction of 5.2% w/v aqueous solution (at 25°C)	7.13	7.1 +/- 0.2

**Chief Chemist**