



Serving Scientific Community  
Since 1985



**TECHNO PHARMCHEM**

An ISO 9001 : 2008 Certified Company

Registered Off. : 101-A, Deep Enclave, Pocket 'D', Ashok Vihar, Phase-III, Delhi-110 052 (India)  
Corporate Off. : 152, Vardhman City Centre, Near Shakti Nagar Underbridge, Delhi-110052 (India)  
Works : Plot No.1022, Modern Industrial Estate, Bahadurgarh-124507, Haryana (India)  
Telephone : 0091 - 11 - 23646422 E-mail : tecpharm@gmail.com  
Website : www.technopharmchem.com

## CATE OF ANALYSIS

**Name of the Sample** : Buffer Tab. pH: 4.0  
**Batch No.** : 227126  
**Date of Mfg.** : Nov 2023  
**Date of Exp** : Oct 2026  
**Register No.** : 2023-24  
**Qty of Sample** : 1 Tab  
**Date of Analysis** : Nov 2023

S.No	Test Parameters	Observed Values	Standard Values
1.	Description	Passes	White uniform tablets
2.	pH of one tablet dissolved in 100 ml distilled water	4.00	4.00 ± 0.05

**Chief Chemist**