



Serving Scientific Community
Since 1985



TECHNO PHARMCHEM

An ISO 9001 : 2008 Certified Company

Registered Off. : 101-A, Deep Enclave, Pocket "D", Ashok Vihar, Phase-III, Delhi-110 052 (India)
Corporate Off. : 152, Vardhman City Centre, Near Shakti Nagar Underbridge, Delhi-110052 (India)
Works : Plot No.1022, Modern Industrial Estate, Bahadurgarh-124507, Haryana (India)
Telephone : 0091 - 11 - 23646422 E-mail : tecpharm@gmail.com
Website : www.technopharmchem.com

CERTIFICATE OF ANALYSIS

Name of the Sample : **Silica Gel 60 - 120 mesh for column chromatography**
Batch No. : **227537**
Date of Mfg. : **May 2024**
Date of Exp : **April 2029**
Register No. : **2024-25**
Qty of Sample : **500 gm**
Date of Analysis : **May 2024**

S.No	Test Parameters	Observed Values	Standard Values
1.	Description	Passes	White Crystalline Powder
2.	pH (10% slurry)	6.95	About 7.0
3.	Mesh particle assay 60-120	Passes	Min.80%
4.	Moisture (LOD)	1.45%	Max.2.0%
5.	Suitability for Chromatography	Passes	To Pass test

Chief Chemist