



Serving Scientific Community  
Since 1985



**TECHNO PHARMCHEM**

An ISO 9001 : 2008 Certified Company

Registered Off. : 101-A, Deep Enclave, Pocket 'D', Ashok Vihar, Phase-III, Delhi-110 052 (India)  
Corporate Off. : 152, Vardhman City Centre, Near Shakti Nagar Underbridge, Delhi-110052 (India)  
Works : Plot No.1022, Modern Industrial Estate, Bahadurgarh-124507, Haryana (India)  
Telephone : 0091 - 11 - 23646422 E-mail : tecpharm@gmail.com  
Website : www.technopharmchem.com

### CERTIFICATE OF ANALYSIS

**Name of the Sample** : PAN Indicator AR  
**Batch No.** : 227686  
**Date of Mfg.** : June 2024  
**Date of Exp** : May 2029  
**Register No.** : 2024-25  
**Qty of Sample** : 25gm  
**Date of Analysis** : June 2024

S.No	Test Parameters	Observed Values	Standard Values
1.	Description	Passes	Dark Orange to brown powder
2.	Solubility 0.1% (95% EtOH)	Passes	Clear Solution
3.	Absorptivity 1% , 1cm at $\lambda$ max.465.6nm	704	> 670
4.	Loss on drying at 110°C	0.84%	< 1.0%
5.	Suitability as Complexometric reagent	Passes	To Pass test

Chief Chemist