



Serving Scientific Community
Since 1985



TECHNO PHARMCHEM

An ISO 9001 : 2008 Certified Company

Registered Off. : 101-A, Deep Enclave, Pocket 'D', Ashok Vihar, Phase-III, Delhi-110 052 (India)
Corporate Off. : 152, Vardhman City Centre, Near Shakti Nagar Underbridge, Delhi-110052 (India)
Works : Plot No.1022, Modern Industrial Estate, Bahadurgarh-124507, Haryana (India)
Telephone : 0091 - 11 - 23646422 E-mail : tecpharm@gmail.com
Website : www.technopharmchem.com

CERTIFICATE OF ANALYSIS

Name of the Sample : Buffer Tab. pH: 4.0
Batch No. : 228059
Date of Mfg. : Oct 2024
Date of Exp : Sep 2027
Register No. : 2024-25
Qty of Sample : 1 Tab
Date of Analysis : Oct 2024

S.No	Test Parameters	Observed Values	Standard Values
1.	Description	Passes	White uniform tablets
2.	pH of one tablet dissolved in 100 ml distilled water	3.99	4.00 ± 0.05

Chief Chemist